

Outpatient Treatment Centers/State ESRD's
Top Ten State Deficiencies
(07-01-08 through 06-30-09)

R9-10-1015.6.b.	Environmental Srvc. And Equipment Standards
R9-10-1003.D.6.j.ii.	Administration
R9-10-1011.1.	Discharge
R9-10-1012.H.	Dialysis Services
R9-10-1014.A.2.d.	Infection Control
R9-10-1003.D.6.g.xviii.	Administration
R9-10-1012.A.5.e.ii.	Dialysis Services
R9-10-1015.1.a.	Environmental Srvc. And Equipment Standards
R9-10-1012.A.4.a.	Dialysis Services
R9-10-1012.A.7.d.	Dialysis Services

0646:

R9-10-1015. Environmental Services and Equipment Standards

An administrator shall ensure that:

R9-10-1015.6. Equipment used at the outpatient treatment center providing dialysis services, medical services, and nursing services:

R9-10-1015.6.b Used according to the manufacturer's recommendations;

0205:

R9-10-1003. Administration

R9-10-1003.D Under the direction of the governing authority of the outpatient treatment center providing dialysis services, medical services, and nursing services, an administrator shall:

R-10-1003.D.6 Ensure that:

R9-10-1003.D.6.j Patient follow-up instructions are:

R9-10-1003.D.6.j.ii If a patient returns to a health care institution, provided orally or in written form, when the patient returns to the health care institution, to the registered nurse responsible for the nursing services provided to the patient at the health care institution or to the individual responsible for the medical services, nursing services, or health-related services provided to the patient at the health care institution;

0433:

R9-10-1011. Discharge

For a patient who received dialysis services, an administrator shall ensure that after the patient's discharge from an outpatient treatment center providing dialysis services, medical services, and nursing services, the nephrologist responsible for the dialysis services provided to the patient documents the patient's discharge in the patient's medical record within 30 days after the patient's discharge and includes:

R9-10-1011.1 A description of the patient's medical condition and the dialysis services provided to the patient

0563:

R9-10-1012. Dialysis Services

R9-10-1012. H A chief clinical officer shall ensure that the quality of water used in dialysis conforms to the guidelines adopted by the Association for the Advancement of Medical Instrumentation in Hemodialysis systems, ANSI/AAMI RD5:2003, incorporated by reference, on file with the Department, and including no future editions or amendments. Copies may be purchased from the Association for the Advancement of Medical Instrumentation, 1110 N. Glebe Road, Suite 220, Arlington, VA 22201-4795.

0603:

R9-10-1014. Infection Control

A. An administrator shall ensure that:

R9-10-1014.2 Policies and procedures are established, documented, and implemented that cover:

R9-10-1014.2.d Sterilization and disinfection of medical equipment and supplies;

0188:

R9-10-1003. Administration

R9-10-1003.D Under the direction of the governing authority of the outpatient treatment center providing dialysis services, medical services, and nursing services, an administrator shall:

R9-10-1003.D.6 Ensure that:

R9-10-1003.D.6.g Policies and procedures are established, documented, and implemented that cover:

R9-10-1003.D.6.g.xviii Dispensation, administration, and disposal of medications and biologicals;

0457:

R9-10-1012. Dialysis Services

A. An administrator of an outpatient treatment center providing dialysis services, medical services, and nursing services shall ensure that for a patient receiving dialysis services:

R9-10-1012.A.5 A registered nurse responsible for the nursing services provided to the patient receiving dialysis services:

R9-10-1012.A.5.e Documents in the patient's medical record:

R9-10-1012.A.5.e.ii Monthly notes related to the patient's progress;

0636:

R9-10-1015. Environmental Services and Equipment Standards

An administrator shall ensure that:

R9-10-1015. 1. An outpatient treatment center providing dialysis services, medical services, and nursing services' premises are:

R9-10-1015. 1.a. Cleaned and disinfected according to the outpatient treatment center providing dialysis services, medical services, and nursing services' policies and procedures to control illness and infection

0448:

R9-10-1012. Dialysis Services

A. An administrator of an outpatient treatment center providing dialysis services, medical services, and nursing services shall ensure that for a patient receiving dialysis services:

R9-10-1012.A.4. The patient's nephrologist or the Nephrologist's designee:

R9-10-1012.A.4.a. Performs a medical history and physical examination on the patient at least once every 12 months from the date of the patient's admission to the outpatient treatment center providing dialysis services, medical services, and nursing services

0463:

A. An administrator of an outpatient treatment center providing dialysis services, medical services, and nursing services shall ensure that for a patient receiving dialysis services:

R9-10-1012.A.7 The patient:

R9-10-1012.A.7.d Receives the dialysis services ordered for the patient by a medical staff member;